

TM Update 2020-005
September 9, 2020

The following Tech Updates were approved by the Board of Trustees, and are effective immediately.

TM 504.3 – BURETTE SPECIFICATION

Test must be performed with a glass Certified or Calibrated “Grade A” burette with a Teflon stopcock.

Add: or certified digital equivalent.